



G MULTI SERVICES

78, Yashob Street, Gobichettipalayam, Erode Dist, Tamil Nadu. Pin: 638 452.

gmultiservicesgobi@gmail.com. Cell : 96597 20684.

Bio Medical Instruments Services

Er. A. Nathul Rasiq DECE., BE (ICE)
Bio Medical Engineer

Calibration Protocol & Report for Medical Equipment

Equipment : MICROSCOPE
MAKE : LABOMED
MODEL : VISION 2000

PROTECT CLASS /TYPE : Risk Class II/CF
S. No. : 211033006
Location : GPHC, ASANUR

S.No	Measurement Parameters		Status	Remarks
1	Electrical Test	(a) Safety Standards (range : 210-240 V) (b) Equipotential ground & Earthing (c) Risk of shock/Fire Hazards, if any (d) Chassis & Insulation Resistance (e) Earth & Chassis Leakage Current (f) Battery conditions & charging status	235 volts 0 to 5volts No Zero ohms 10mA —	Stabilizer/Ups With in Ranges Well
2	Physical Condition Test	(a) Cabinet & visible damages, if any (b) Power Supply unit, halogen lamps, Power cord accessories (c) Risk of mechanical damages, if any	No OK No	Checked ok
3	Performance Test	(a) Halogen lamps intensity (b) Intensity test taken by lux meter for a particular lamp & results (c) Measurement of lux/Lens range (d) Lamp Life time	Normal OK OK OK	Fine Brightness Perfect 20000 hrs Apprx.
4	Calibration Test	(a) LED lamps intensity (b) Intensity test taken by lux meter for a particular lamp & results (c) Measurement of lux/Lenses cal (d) Lamp life time	OK Done 1200 Lux OK	In range Peak In range In range
Formulated By: Edition Research Laboratory		Issued By:		User Department:
Endorsed By: Research team head				
Reviewed On: 22.09.23 Next Review : 21.09.24				