

	<p align="center"><b>ROOTS INDUSTRIES INDIA LIMITED</b>  <b>ROOTS METROLOGY &amp; TESTING LABORATORY</b>  R.K.G Industrial Estate, Ganapathy, Coimbatore-641  006,India, Phone :9003432356</p>	 <p align="center">NABL/ILAC/024 NABL ACCREDITATION  CERTIFICATE NO CC-2752</p>
----------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------------------------

**CALIBRATION CERTIFICATE**

ULR Number: CC275219000083790F

Name and address of the customer The Medical officer Government urban Primery Health Centre pattukottai–thanjavur( dist)	Certificate no : Rmtl/01/319107849/4 Date of calibration :03/08/2023 Suggested calibration due : 02/08/2024 Customer po .reference.& date:
-----------------------------------------------------------------------------------------------------------------------------------	-----------------------------------------------------------------------------------------------------------------------------------------------------

**DESCRIPTION OF DUC**

Name Of The Duc: pipette Make : THERMOFISHIER Model : finnpipette. f3 Sl.No : NA Range/Size : 100-1000 micro litter	Receiving Condition Of Duc :Good Ref. Document :As Per Manufactures Specification Environmental Condition :25.3 Deg & 54.3% Rh Uncertainty Of Measurement :As Per Report Location :lab
---------------------------------------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------

**THE MASTERS ARE TRACEABLE TO NATIONAL/INTERNATIONAL STANDARDS AS PER ISO/IEC 17025.**

**DETAILS OF MASTER USED**

Standard Instrument Used	Make/ Model	Id No	Traceable To/Certificate No	Validity
Multi Product Calibrator	Fluke	Rml-262	Wavecal,Blr/Cs-18/01-01	1-6-2024
Micro balance	mettle toledo/XP26PC	QP-01	B115129039/CC-2523	1-6-2024
Universal Measuring Instrument With Sensor For Envirmental Parameter	ALMEMO/2590-4AS	QP-10	H19070544/CC-2052	1-6-2024

**CALIBRATION FIELD: WEIGHT Calibration**

Sl.no	Range	Master readings micro litter	Calibrated values micro litter	Error observed micro litter	Uncertainty
1	10-100 micro litter	100	99.9	0.1	0.055
2	10-100 micro litter	500	499.9	0.1	0.055
3	10-100 micro litter	800	799.9	0.1	0.055
4	10-100 micro litter	900	899.9	0.1	0.055
5	10-100 micro litter	1000	999.9	0.1	0.055

remarks

uncertainty estimated for a confidence level of 95% at the k factor=2  
the reported results are valid only for the condition of the inst./gauge received at the time of and under the condition of the calibration  
DUC –device under calibration  
Calibration label /status : Calibration result pass & fail are indicated

			This Report Refers Only To The Sample Submitted And May Not Be Reproduces In Full Or In Part, Without Written Permission From Roots Metrology& Testing Laboratory, Coimbatore, cell:9585840192
-------------------------------------------------------------------------------------	-------------------------------------------------------------------------------------	--------------------------------------------------------------------------------------	------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------