



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - AUGUST 2023



PC-1024

Lab Name Govt. PHC, Kodangipatti

Lab No 12304

Constituent Group Chemistry III

Date of Result Entered : 19/08/2023

PT item Lyophilized human serum based

Date of Report Published : 01/09/2023

SLNo	Analyte	Method/ Principle Name	Analyzer Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	130 mg/dL	-5.06	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	2.4 mg/dL	5.33	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	125 mg/dL	2.42	0.22

SDI Range

Interpretation

Within -1.00 to +1.00

Excellent.

Within ±1.01 to ±2.00

Good.

Within ±2.01 to ±2.99

Accept with caution. Warning Signal.

beyond ±3.0

Unacceptable performance. Action Signal.