

VIEW LAB MONTHLY SUMMARY

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

Lab Name Govt. PHC, Avanipur

Lab No 13089

Month June

Year 2023

Constituent Group Chemistry III

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 19/06/2023

Date of Report Published : 04/07/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|-------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1823 | 286.63 | 5.17 | 14.82 | 298 mg/dL | 0.77 | 0.69 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1848 | 2.57 | 6.64 | 0.17 | 2.8 mg/dL | 1.34 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1804 | 113.73 | 4.46 | 5.07 | 108 mg/dL | -1.13 | 0.24 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|----------------|-----------------|--------------------|--------------------------|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |


Medical Officer
Govt. Primary Health Centre
Avanipur

VIEW LAB MONTHLY SUMMARY

Details About Robust Analysis

Detail About Monthly Summary

Detail about SDI

All Analyser Result

Print

Date of Result Entered : 19/05/2023

Date of Report Published : 05/06/2023

Lab Name: Govt PHC, Avaniapur
 Lab No: 1000
 Month: May
 Year: 2023
 Department: Chemistry III

Click on the analyte to view Graphical Data

| Sl No | Analyte | Method / Principle Name | Analyser Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|-----------------|--------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1834 | 180.88 | 5.19 | 9.39 | 165 mg/dL | -1.69 | 0.44 |
| 2 | CREATININE III | Jaffe Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1834 | 1.76 | 7.48 | 0.13 | 1.6 mg/dL | -1.21 | 0.01 |
| 3 | CHOLESTEROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1798 | 114.60 | 5.08 | 5.83 | 120 mg/dL | 0.93 | 0.27 |

| SDI Range | Interpretation |
|---------------------------------|--|
| Within ± 1.00 to ± 1.00 | Excellent. |
| Within ± 1.01 to ± 2.00 | Good. |
| Within ± 2.01 to ± 2.99 | Accept with caution. Warning Signal. |
| Beyond ± 3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|----------------|-----------------|--------------------|--------------------------|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Verify Summary | Change Password | Home | Sign-out |

12.05.23

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Avanipur
Lab No 13089
Month April
Year 2023
Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

Click on the analyte to view Graphical Data

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/04/2023

Date of Report Published : 06/05/2023

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No of Participants | DV | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1832 | 119.45 | 6.04 | 7.21 | 110 mg/dL | -1.31 | 0.34 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1820 | 2.16 | 8.09 | 0.18 | 2.3 mg/dL | 0.80 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1750 | 111.61 | 5.19 | 5.79 | 120 mg/dL | 1.45 | 0.28 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |

R. Deepu.