

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Avanipur

[Details About Robust Analysis](#)

Lab No 13089

[Detail About Monthly Summary](#)

Month September

[Detail about SDI](#)

Year 2022

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/09/2022


Date of Report Published : 05/10/2022

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1846	223.49	6.67	14.90	226 mg/dL	0.17	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1834	3.49	9.05	0.32	3.2 mg/dL	-0.92	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	112.08	5.53	6.20	110 mg/dL	-0.34	0.29

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out


Medical Officer
Govt. Primary Health Centre
Avanipur