

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13381 Page No: 1 of 1

Customer Name & Address		SRF No.	SRF/23/00735-0004
GOVERNMENT PRIMARY HEALTH CENTRE, VADAKKUPALAYAM-639004,KARUR DISTRICT.		SRF Date	07-10-2023
		Date of Receipt	06-10-2023
		Date of Calibration	07-10-2023
		Due Date for Calibration	06-10-2024
		Issue Date	07-10-2023

Details of Unit Under Calibration			
Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	FINNPIPETTE F3
Resolution	1µl	Material	PVC
Serial Number	RW11887	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY


Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08


Standard Used						
Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023


Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
		1	2	3	4	5				
1	100	99.87	99.88	99.85	99.84	99.86	99.85	-0.15	0.02	0.47
		99.82	99.83	99.84	99.85	99.87				
2	500	499.95	499.96	499.95	499.96	499.97	499.96	-0.04	0.01	0.47
		499.96	499.97	499.98	499.98	499.96				
3	1000	999.80	999.80	999.81	999.82	999.83	999.81	-0.19	0.01	0.47
		999.80	999.82	999.83	999.80	999.81				

Remarks

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,

 (Calibration Engineer)
 MADHIBAN



Authorised by:

 (Quality Manager/Chief Executive)
 C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.