



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name Govt. UPHC, Rameswaram

Lab No

12571

Constituent Group Chemistry III

Date of Result Entered :

20/04/2023

PT Item Lyophilized human serum based

Date of Report Published :

06/05/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Your Value	SDI	U
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	0.04	7.21	116 mg/dL	-0.48	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	112 mg/dL	0.07	0.28
SDI Range										
				Within -1.00 to +1.00						
				Excellent.						
				Within +1.01 to +2.00						
				Good.						
				Within +2.01 to +2.99						
				Accept with caution. Warning Signal.						
				Beyond +3.0						
				Unacceptable performance. Action Signal.						

LAB ADDRESS :

Govt UPHC, Rameswaram
Ponhampuli K P KOVIL
Uphc Rameswaram
TAMILNADU623529

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Dr. Pamela Christodoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Pamela Christodoss

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential