
AMS Healthcare

(Erstwhile : Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony, Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610



GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

Customer	M/S. Government Primary Health Centre,	Certificate No:	AMS/LAB/289/4924
	Sathangudi -625706. Madurai Dist.	Issuing Date:	30/09/2023
Address:		Page No:	1 of 1

Details of Device under Calibration (DUC):

DUC:	Micropipette	Id. No:	PHC/LAB/VMP/003
Make:	Thermo Scientific	Department:	Laboratory
Model:	Finnpipette'F3	Calibrated On:	30/09/2023
SI. No:	RW13097	Calibration	
Type:	Variable	Due Suggested:	29/09/2024
Range &	100 to 1000µl	DUC Condition	
Least Count:		on Receipt:	ОК

Traceability Details:

Sl.No.	Test Equipment	Make	Certificate Number	Valid Up to
1	Electronic Weighing Scale	EWC/SF400C	PC/01/23/ATL/0512-02	31/01/2024
2	Thermo Hygrometer	Beetech/088	PC/01/23/ATL/0210-02	31/01/2024
Room T	emperature: 25°C±2 Room	h Humidity: 30-60%	% (At the time of Calibration	

Room Temperature: 25°C±2 Test ResultsVolume µl:

SL.	DUC Reading	Standard Reading (µl)		Correction
No.	(µl)	Trial 1	Trial 2	(µl)
1	100	100.0	100.0	+00.0
2	200	200.0	200.0	+00.0
3	500	500.0	500.0	+00.0
4	800	800.0	800.0	+00.0
5	1000	1000.0	1000.0	+00.0

Allowable Tolerance: User Defined



Note:

1. Calibration Points were selected as per customer request.

2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of lo

3. Results Reported are valid at the time of and under the stated conditions of measurements.

Calibrated By:

(Calibration Engineer)



Approved by: (Technical Manager)

94434 74610, 93455 56148

