AMS Healthcare

(Erstwhile : Ace Medical Solutions (India) Pvt. Ltd.)

An ISO 9001-2015 Certified Company

No. 770/1, Jothi Complex 2nd Floor, Alagar Kovil Main Road, Surveyor Colony, Madurai - 625 007, Tamilnadu, India Mobile : +91 94434 74610

GSTIN. 33AOSPA9153Q1ZA • DL No. TN-12-20B -00398, TN-12-21B-00398

CALIBRATION CERTIFICATE

	M/S. Government Primary Health Centre,	Certificate No:	AMS/LAB/289/4923	
Customer	Sathangudi -625706. Madurai Dist.	Issuing Date:	30/09/2023	
Tunite et	Sathangudi -025700. Madurai Dist.	Page No:	1 of 1	
Address:			4	_

Details of Device under Calibration (DUC):

DUC	Micropipette	1d. No:	PHC/LAB/VMP/002
DUC:		Department:	Laboratory
Make:	Thermo Scientific	Calibrated On:	30/09/2023
Model:	Finnpipette'F3		
SI. No:	PW00636	Calibration Due	29/09/2024
Type:	Variable	Suggested:	29/09/2024
Range &	10 to 100µl	DUC Condition on	OV
Least Count:		Receipt:	ОК

Traceability Details:

SI.No.		Make	Certificate Number	Valid Up to
31.140.	Test Equipment		PC/01/23/ATL/0512-02	31/01/2024
1	Electronic Weighing Scale	Encidi tore	PC/01/23/ATL/0210-02	31/01/2024
2	Thermo Hygrometer	Beetech/088		f Calibration)

(At the time of Calibration) Room Humidity: 30-60% Room Temperature: 25°C±2

Test ResultsVolume µl:

SL.	DUC Reading	Standard Reading (µl)		Correction
No.	(µl)	Trial 1	Trial 2	(µl)
1	10	10.0	10.0	+00.0
1	20	20.0	20.0	+00.0
2		50.0	50.0	+00.0
3	50	100.0	100.0	+00.0
4	100	100.0	100.0	

Allowable Tolerance: User Defined



Note:

1. Calibration Points were selected as per customer request.

2. Each reading was taken by Repeatability Test(i.e.) each reading in a row is a mean of 10 r

the stated conditions of measurements. 3. Results Reported are valid at the time of and

Calibrated By:

(Calibration Engineer)



aceagm@gmail.com

Du amshealthcaremdu@gmail.com

Approved by: (Technical Manager)



94434 74610, 93455 56148