

Report No : HBMS/CAL/23-24

CALIBRATION REPORT

1.1. CUSTOMER DETAILS:		Calibration Date: 29.09.2023
		Calibration valid: 28.03.2024
Name and Address of the organization		M/s, GOVT. PRIMARY HEALTH CENTRE, PULLIPALYAM, SANKARI, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		
i	Manufactured by	MICRO PIPETTE
ii	Model	THERMOSCIENTIFIC
iii	Value	FINNPIPETTE
iv	Serial No	100µl-1000µl
		GPHC/PULM/PIP/005


Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	1000.08
5	100.00	499.94	1000.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,


 (Authorized signatory)

MD Reg No: TN/CBE/MD42/00016
 GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106
 88833 15662

No.42/1, Navaindia Road, Peela
 CBE - 04, hbmedicalsystem@gmail.com