



## CALIBRATION CERTIFICATE

CUSTOMER ADDRESS : M/s. THE MEDICAL OFFICER GOVT. PRIMARY HEALTH CENTRE RAMANATHAPURAM  COIMBATORE Phone No.:	
UNIQUE LAB REPORT NUMBER : ULR-CC231723000000304F	DISCIPLINE : MECHANICAL CALIBRATION
NAME OF EQUIPMENT : MICROPIPETTE	PARAMETER : VOLUME
CERTIFICATE NO : TPMCC230156015	DATE OF ISSUE : 23-01-2023
RECEIPT NO : 11516	TEMPERATURE : 23.6 °C
RECEIPT DATE : 19-01-2023	HUMIDITY : 52 %
JOB CARD NO : TPM410012354438	AIR PRESSURE : 962 hpa
JOB CARD DATE : 19-01-2023	MEDIUM OF WEIGHING : GRADE III WATER
BRAND : FINNPIPETTE - F3	DENSITY OF MEDIUM : APPROXIMATELY 1
RANGE : 10 - 100 µl	DENSITY CORRECTION FACTOR [Z] : 1.0033 µl/mg
SERIAL NO : RW01615	CALIBRATION METHOD : ISO 8655:2002 PART - 6
TIPS USED : 200 µl	WEIGHING BALANCE : SEMI MICRO BALANCE
SCOPE OF ACCREDITATION : VOLUME - MICROPIPETTE ( 1 µl - 10000 µl )	

### DETAILS OF MASTER EQUIPMENTS USED FOR CALIBRATION

MASTER EQUIPMENT	: SARTORIUS
SL.NO	: 19402655
RESOLUTION	: 0.00001 g upto 80 g & above is 0.0001 g
CALIBRATED BY	: THE PIPETEMANN(MASS),COIMBATORE
CERTIFICATE NUMBER	: TPMCCM22050373
CALIBRATION VALIDITY	: 03-05-2023
TRACEABILITY	: NATIONAL STANDARDS