



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Erasakkanaicknoor	Details About Robust Analysis
Lab No	12704	Detail About Monthly Summary
Month	March	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1878	91.48	7.46	6.82	93 mg/dL	0.22	0.31
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	175	55.20	11.70	6.46	54 mg/dL	-0.19	0.98
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1851	1.33	9.08	0.12	1.3 mg/dL	-0.25	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	43	4.73	7.17	0.34	4.6 mg/dL	-0.38	0.10
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1761	110.93	5.23	5.80	115 mg/dL	0.70	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Erasakkanaicknoor
 Lab No 12704
 Month September
 Year 2023
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 20/09/2023

Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	303 mg/dL	-0.52	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	187	35.19	12.73	4.48	29 mg/dL	-1.38	0.66
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.2 mg/dL	-0.86	0.00
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	35	6.72	4.36	0.29	6.7 mg/dL	-0.07	0.10
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	113 mg/dL	-0.14	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name	Govt. PHC, Erasakkanaicknoor	Lab No	12704
Constituent Group	Chemistry III	Date of Result Entered :	19/08/2023
PT Item	Lyophilized human serum based	Date of Report Published :	01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1841	175.58	5.13	9.00	180 mg/dL	0.49	0.42
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	174	28.29	15.20	4.30	35 mg/dL	1.56	0.65
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.76	6.77	0.12	1.73 mg/dL	-0.25	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	40	5.43	10.18	0.55	5.3 mg/dL	-0.24	0.17
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1764	113.80	4.07	4.64	115 mg/dL	0.26	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. PHC, Erasakkanaicknoor
GOVT.HR.SE.SCHOOL, NEAR
Erasakkanaicknoor
TAMILNADU625515

Coordinator Contact Details:
Email:clinq@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****



VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Erasakkanaicknoor [Details About Robust Analysis](#)
Lab No 12704 [Detail About Monthly Summary](#)
Month July [Detail about SDI](#)
Year 2023
Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 19/07/2023
Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	120 mg/dL	1.76	0.28
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	184	51.41	11.14	5.73	47 mg/dL	-0.77	0.84
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.9 mg/dL	0.20	0.02
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	30	5.35	9.89	0.53	5.9 mg/dL	1.04	0.19
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	114 mg/dL	0.21	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Erasakkanaicknoor	Details About Robust Analysis
Lab No	12704	Detail About Monthly Summary
Month	June	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/06/2023

Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	280 mg/dL	-0.45	0.69
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	192	32.69	12.96	4.24	30.87 mg/dL	-0.43	0.61
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.23 mg/dL	-1.99	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	34	6.25	4.51	0.28	6.2 mg/dL	-0.18	0.10
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	116.51 mg/dL	0.55	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Erasakkanaicknoor	Details About Robust Analysis
Lab No	12704	Detail About Monthly Summary
Month	May	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/05/2023

Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1834	180.88	5.19	9.39	175 mg/dL	-0.63	0.44
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	163	27.91	13.39	3.74	20.37 mg/dL	-2.02	0.59
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1854	1.76	7.48	0.13	1.77 mg/dL	0.08	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	40	4.98	10.94	0.55	4.61 mg/dL	-0.68	0.17
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	114.60	5.08	5.83	117 mg/dL	0.41	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out



VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Erasakkanaicknoor	Details About Robust Analysis
Lab No	12704	Detail About Monthly Summary
Month	April	Detail about SDI
Year	2023	
Constituent Group	Chemistry III	

[Click on the analyte to view Graphical Data](#)

Date of Result Entered : 20/04/2023

Date of Report Published : 06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	119.45	6.04	7.21	119 mg/dL	-0.06	0.34
2	UREA III	Urease UV / GLDH III	Robonik (Semi automatic / Fully Automatic)	169	42.32	10.68	4.52	42 mg/dL	-0.07	0.70
3	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.14 mg/dL	-0.11	0.01
4	URIC ACID III	Enzymatic / Uricase Colorimetric III	Robonik (Semi automatic / Fully Automatic)	38	4.58	9.59	0.44	4.9 mg/dL	0.73	0.14
5	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	5.19	5.79	114 mg/dL	0.41	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links			
Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out