

THE PIPETEMANN

NABL ACCREDITED CALIBRATION LABORATORY
ADDING ACCURACY TO YOUR EQUIPMENT

MASS - VOLUME - THERMAL - SPEED - PRESSURE - TIMER - ELECTRO TECHNICAL



CC - 2317

CALIBRATION CERTIFICATE

CUSTOMER ADDRESS : M/s. THE MEDICAL OFFICER GOVT. PRIMARY HEALTH CENTRE RAMANATHAPURAM COIMBATORE Phone No.:	
UNIQUE LAB REPORT NUMBER : ULR-CC231723000000305F	DISCIPLINE : MECHANICAL CALIBRATION
NAME OF EQUIPMENT : MICROPIPETTE	PARAMETER : VOLUME
CERTIFICATE NO : TPMCC230156016	DATE OF ISSUE : 23-01-2023
RECEIPT NO : 11516	TEMPERATURE : 23.6 °C
RECEIPT DATE : 19-01-2023	HUMIDITY : 52 %
JOB CARD NO : TPM410012354439	AIR PRESSURE : 962 hpa
JOB CARD DATE : 19-01-2023	MEDIUM OF WEIGHING : GRADE III WATER
BRAND : FINNPIPETTE - F3	DENSITY OF MEDIUM : APPROXIMATELY 1
RANGE : 100 - 1000 µl	DENSITY CORRECTION FACTOR [Z] : 1.0033 µl/mg
SERIAL NO : RW02714	CALIBRATION METHOD : ISO 8655:2002 PART - 6
TIPS USED : 1000 µl	WEIGHING BALANCE : SEMI MICRO BALANCE
SCOPE OF ACCREDITATION : VOLUME - MICROPIPETTE (1 µl - 10000 µl)	

DETAILS OF MASTER EQUIPMENTS USED FOR CALIBRATION

MASTER EQUIPMENT	: SHIMADZU
SL.NO	: D450020184
RESOLUTION	: 0.00001g upto 82g & above is 0.0001g
CALIBRATED BY	: THE PIPETEMANN (MASS), COIMBATORE
CERTIFICATE NUMBER	: TPMCCM22020355
CALIBRATION VALIDITY	: 03-02-2023
TRACEABILITY	: NATIONAL STANDARDS