



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



PC-1024

Lab Name	Govt. Block PHC, Kolathur	Lab No	7496
Constituent Group	Chemistry III	Date of Result Entered :	20/04/2023
PT item	Lyophilized human serum based	Date of Report Published :	06/05/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	134 mg/dL	2.02	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.1 mg/dL	-0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	106 mg/dL	-0.97	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. Block PHC, Kolathur  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
Kolathur  
TAMILNADU636303

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - MAY 2023



Lab Name Govt. Block PHC, Kolathur Lab No 7496  
Constituent Group Chemistry III Date of Result Entered : 20/05/2023  
PT item Lyophilized human serum based Date of Report Published : 05/06/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1834	180.88	5.19	9.39	176 mg/dL	-0.52	0.44
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1854	1.76	7.48	0.13	1.7 mg/dL	-0.45	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1798	114.60	5.08	5.83	116 mg/dL	0.24	0.27

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. Block PHC, Kolathur  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
Kolathur  
TAMILNADU636303

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JUNE 2023



Lab Name Govt. Block PHC, Kolathur Lab No 7496  
Constituent Group Chemistry III Date of Result Entered : 20/06/2023  
PT item Lyophilized human serum based Date of Report Published : 04/07/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1823	286.63	5.17	14.82	281 mg/dL	-0.38	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1848	2.57	6.64	0.17	2.4 mg/dL	-0.99	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1804	113.73	4.46	5.07	109 mg/dL	-0.93	0.24

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. Block PHC, Kolathur  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
Kolathur  
TAMILNADU636303

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - AUGUST 2023



Lab Name Govt. Block PHC, Kolathur Lab No 7496  
Constituent Group Chemistry III Date of Result Entered : 19/08/2023  
PT item Lyophilized human serum based Date of Report Published : 01/09/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1841	175.58	5.13	9.00	191 mg/dL	1.71	0.42
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.76	6.77	0.12	1.8 mg/dL	0.34	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1764	113.80	4.07	4.64	122 mg/dL	1.77	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. Block PHC, Kolathur  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
Kolathur  
TAMILNADU636303

Coordinator Contact Details:  
Email:clinq@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. Block PHC, Kolathur Lab No 7496  
Constituent Group Chemistry III Date of Result Entered : 20/09/2023  
PT item Lyophilized human serum based Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1870	310.86	4.82	14.98	302 mg/dL	-0.59	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1852	1.29	8.14	0.10	1.3 mg/dL	0.10	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	113.62	3.80	4.32	117 mg/dL	0.78	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. Block PHC, Kolathur  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
Kolathur  
TAMILNADU636303

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - JULY 2023



Lab Name Govt. Block PHC, Kolathur Lab No 7496  
Constituent Group Chemistry III Date of Result Entered : 20/07/2023  
PT item Lyophilized human serum based Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1876	109.35	5.55	6.07	110 mg/dL	0.11	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1850	5.83	6.00	0.35	5.8 mg/dL	-0.09	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1776	113.00	4.12	4.66	117 mg/dL	0.86	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. Block PHC, Kolathur  
THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, SALEM DISTRICT.  
Kolathur  
TAMILNADU636303

Coordinator Contact Details:  
Email:clinqc@cmcvellore.ac.in  
Contact Number: 0416-2283102

*Pamela Christudoss*

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components

\*\*\*\*\* End of Report \*\*\*\*\*