

<b>CALIBRATION REPORT</b>	
Report No: HBMS/CAL/23-24	
Calibration Date: 04.10.2023	
Calibration Valid: 03.04.2024	
<b>1.1. CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, URBAN PRIMARY HEALTH CENTRE, NARASINGAPURAM, SALEM.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	PIPETTE
I	Manufactured by
	---
II	Model
	10µl
III	Serial No
	UPHC/NRM/PIP/001
IV	Location
	LABORATORY

**Calibration standards used :**

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

**Details of the machine:** machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)
	10µl
1	10.05
2	10.06
3	10.01
4	9.97
5	10.01
6	10.02

**For Heart Beat Medical Systems**

**(Authorized signatory)**

