

| CALIBRATION REPORT | |
|---|---|
| Report No: HBMS/CAL/23-24 | |
| Calibration Date: 04.10.2023 | |
| Calibration Valid: 03.04.2024 | |
| 1.1. CUSTOMER DETAILS: | |
| Name and Address of the organization | M/s, URBAN PRIMARY HEALTH CENTRE, NARASINGAPURAM, SALEM. |
| 1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT): | |
| EQUIPMENT NAME | PIPETTE |
| I | Manufactured by |
| II | Model |
| III | Value |
| IV | Serial No |

Calibration standards used :

| Master instrument | Model | Serial no | Date of calibration | Calibration due on |
|--------------------------|---------|------------|---------------------|--------------------|
| Electrical compact scale | SF-400C | SF400-2240 | 07-OCT-2022 | 06-OCT-2023 |

Details of the machine: machine calibrated and checked output.

| Number of measurement | Channel 1 (Volume) |
|-----------------------|--------------------|
| | 1000µl |
| 1 | 1000.05 |
| 2 | 1000.06 |
| 3 | 1000.01 |
| 4 | 999.97 |
| 5 | 1000.01 |
| 6 | 1000.02 |

For Heart Beat Medical Systems,

(Authorized signatory)