

CALIBRATION CERTIFICATE

Customer name And Address	Government Primary Health Center, Elampillai.	Certificate No.	SB/2022/0142
		Date of calibration	24.09.2022
		Valid up to	23.09.2023
		Date of issue	26.09.2022
		Calibrated at	On site

<u>Details of Test Instrument</u>			
Instrument name	Semi Auto Analyzer	S.no	AT0771010RBK
Make	Robonik	Model	Preitest Touch
Range/Size	-	Location	Blood Bank Storage Center
Least Count	-	Accuracy	As per manual
I.D No.	-	Visual Inspection	Ok

<u>Details of standards used</u>	
Name	-
Certificate No.	-
Valid up to	-
Traceability	-

Environmental Condition	Temperature	30 ⁰ C		
	Relative Humidity	60%		

Calibration Results

- > Checked the analyzer and internal calibration done.
- > The machine found working in a good condition.


Remarks:

- ❖ The above UUC was calibrated at On-site
- ❖ UUC is defined as Unit under Calibration.
- ❖ This result of calibration refers to only to the particular item submitted for calibration.
- ❖ The above results are valid at the time of and under the stated conditions measurement.

Note: This machine is working fine and measured values are within the range



மருத்துவ அலுவலர்
அரசு மேம்படுத்தப்பட்ட ஆரம்ப
சுகாதார நிலையம்,
இளம்பிள்ளை - 637 502 சேலம் (T.M)



Chief Technical Coordinator
Karthick Madheswaran B.E, MBA.

#236, Kamaraj Nagar Colony, Salem-636014.
Mobile: +91 9791849566, Email: standardbiomedical93@gmail.com