



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. Block PHC, Kiliyanur Lab No 7281
Constituent Group Chemistry III Date of Result Entered : 18/07/2023
PT Item Lyophilized human serum based Date of Report Published : 07/08/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	Participants		Target Value	SDI	SI
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	107 mg/dL	-0.39	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	6.3 mg/dL	1.34	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	111 mg/dL	-6.43	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Kiliyanur

THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, VILLUPURAM DISTRICT.

Kiliyanur

TAMILNADU604102

Coordinator Contact Details:

Email: clinqc@cmcvellore.ac.in

Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss

CMC EQAS Coordinator

Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****

P. Devi
MEDICAL OFFICER
GOVT PRIMARY HEALTH CENTRE
KILIYANUR - 604 102
VILLUPURAM DISTRICT