

	CAL	IBRATIO	N CERT	TFICA	<u>ΓΕ</u>		
CERTIFICATE NO: SBS/CL/23/13448			MEDICAL DEVICES			Page No:1 of 1	
Issue Date	ssue Date		07-10-2023				
SRF No & Date		SRI	SRF/23/00751-0001 & 07-10-2023				
Receipt Date			07-10-2023				
Calibration Date			07-10-2023				
Calibration Due			06-10-2024				
Customer Name & Add	ress						
GOVERNMENT PRIM	IARY HEALTH CENTR	E,					
KAVALKARANPATTY	KARUR DISTRICT.						
		Details of Device	Under Calibrat	ion (DUC)			
Description :	SEMI AUTO ANALYZER		Make & Model :		ROBONIK & PRIETEST TOUCH		
Range :	MULTI	Sr. I	Sr. No :		AT2180211RBK		
Resolution :	MULTI	Ider	tification No	:	NA		
DUC Condition :	SATISFACTORY	Loca	ation	:	LAB		
	Environ	mental Conditions	& Calibration	Procedure D	etails		
Environmental Details Temperature:25.6°C		25.6°C	Relative Humidity 52%		52% RH		
Calibration Procedure No SBS/CP/MD/20		20	Calibration done at ONSITE				
		Reference	Standards Deta	ails		7	
S.No Description Make/ SI No:		Make/ SI No:		Certificate No		Validity	
1 Electrical Safety Analyser Rigel Medical		Rigel Medical & 44	& 44L-1059 M-230		M-230809-16-4		

ELECTRICAL SAFETY

_	_	_	_		_
R	E	SI	J	LT	S

SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)
>20MΩ	78	13.92
Earth Leakage	Measured values in μA	Uncertainty in % (±)
<5000µAfor B,BF,CF	214	8.9
Enclosure Leakage	Measured values in μA	Uncertainty in % (±)
<500µAfor B,BF,CF	204	9.1
	Insulation Resisitance >20MΩ Earth Leakage <5000μAfor B,BF,CF Enclosure Leakage	Insulation Resisitance

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

(Calibration Engineer) K.SATHYAMOORTHY



Technical Manager	r
C.SHANMUGARA	J

. . .

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS