

				<u>CALIE</u>	BRAT	TION	CERT	FICATE			
CERTIFICATE NO:		SBS/CL/23/13449				MECHANICAL		Pa	Page No:1 of 1		
Issue [Date					07-10-20)23				
SRF No & Date					SRF/23/00751-0002 & 07-10-2023						
Receipt Date						07-10-2023					
Calibration Date						07-10-2023					
Calibration Due						06-10-2024					
Custo	mer Name &	& Addr	ess								
GOVE	RNMENT P	RIMAR	RY HEALTH	CENTRE,							
KAVAL	KARANPA	TTY,KA	ARUR DISTR	RICT.							
en of Service Mark				Deta	ails of D	evice Un	der Calibrati	ion (DUC)			
Description		:	CENTRIFUGE			Make & Model :		LABOTECH & BDI-152			
Range		:	3500 RPM			Sr. No :		NA			
Resolution		:	1 RPM			Identification No :		NA			
DUC Condition		:	Satisfactory			Location : LAB					
			Envi	ronmental C	ondition	ns & Star	dard Opera	ting Procedure	e Details		
Environmental Details Temperature: 24.3				e: 24.3 °C	C Relative Humidity		nidity	54 % Rh			
Calibration Procedure No SBS/CP/ML/04			/04	Calibration done at		ONSITE					
				<u> </u>	Refer	ence Sta	indards Deta	ails			
S.No Description Make/				Make/ S	SI No:		Certificate No		Validity		
1	Digital Tachometer Lutron				& Q638380		JRPM-CCTR-	A&S-2022-0036	03-10-2023		
CALIB	RATION RE	SULT	<u>s</u>								
S.No	S.No CALIBRATION READINGS STANDARD INST			INSTRU	JMENTS	DEVIATION		EXPANDED UN	ICERTAINTY (±)		

S.No	CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)	
	RPM	RPM	RPM	%	
1	1000	999.7	0.3	4.2	
2	2000	1997	3	4.2	
3	3000	2997	3	4.2	

REMARKS

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.

2. The user should determine the suitability of the instrument for its intended use.

3. The recalibration interval should be determined on the user requirement.

4. The results stated in this certificate relate only to the item calibrated.

5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.

6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) K.SATHYAMOORTHY



ang Quality Manager (C.SIVABALAN)

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.

Authorised Signatory