



## CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name Govt PHC, Pasuvanthani  
Constituent Group Chemistry III  
PT Item Lysophilized human serum based

Lab No 10636  
Date of Result Entered: 20/04/2023  
Date of Report Published: 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No. of Participants	DV	CV	Participants 90	Your Value	SD	U
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1832	119.45	6.04	7.21	110 mg/dL	-1.31	0.34
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1820	2.16	8.09	0.18	2.2 mg/dL	0.23	0.61
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1750	111.61	5.19	5.79	109 mg/dL	-0.45	0.28
<b>SDI Range</b>				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within 1.01 to 2.00				Good.						
Within 2.01 to 2.99				Accept with caution, Warning Signal.						
Beyond 3.0				Unacceptable performance, Action Signal.						

LAB ADDRESS :  
Govt PHC, Pasuvanthani  
CLINICAL LABORATORY , TUTICORIN District ,  
Pasuvanthani  
TAMILNADU628718

Coordinator Contact Details :  
Email:cdmg@cmconclorac.in  
Contact Number: 0416-2283102

*Ravida Chidambaram*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential  
CMC EQAS does not sub contract any components  
\*\*\*\*\* End of Report \*\*\*\*\*

*[Signature]*  
MEDICAL OFFICER  
GOVT. PRIMARY HEALTH CENTRE  
PASUVANTHANI  
KOVELPATTI - HUD