

# HOPE ULTRA SYSTEMS

No: 2/79, MAIN ROAD, MAMSAPURAM (PO),  
SIVAKASI, VIRUDHUNAGAR-626124.  
E-mail : [hopeultrasystems@gmail.com](mailto:hopeultrasystems@gmail.com)

GSTIN. 33DLMPK3057D1Z9  
Mobile: +919487402761, +919715223192.

## CALIBRATION CERTIFICATE

Requested By

**GPHC, VADAMALAPURAM.**

### Scope:

|                |                |                     |                   |
|----------------|----------------|---------------------|-------------------|
| Equipment Name | : Micropipette | Asset no            | : PHC/VMP-LAB/004 |
| Make           | : MICROLUX     | Calibration Done at | : customer place  |
| Model          | : 5-50µl       | Calibration Date    | : 31/08/2023      |
| Sl no          | : ....         | Next call due on    | : 30/08/2024      |

### TEST RESULTS Volume µl

| Sl. No. | DUC reading(µl) | Standard Reading (µl) |         | Correction (µl) |
|---------|-----------------|-----------------------|---------|-----------------|
|         |                 | Trail 1               | Trail 2 |                 |
| 01      | 05              | 05.0                  | 05.0    | +0              |
| 02      | 10              | 10.0                  | 10.0    | +0              |
| 03      | 20              | 20.0                  | 20.0    | +0              |
| 04      | 40              | 40.00                 | 40.0    | +0              |
| 05      | 50              | 50.0                  | 50.0    | +0              |

Error Acceptance : Not Applicable (user defined)

Equipment Status : Working in Good condition

Materials used for testing and calibration:

3. Electronic weighing scale(EWC/SF400C)

Note:

7. Calibration points were selected as per customer request.
8. Each reading was taken by repeatability.
9. Results reported are valid at the time of and under the stated conditions of measurements.

Hope Ultra Systems



Authorized signatory