

## CALIBRATION CERTIFICATE

Certificate No: VCPL/CC/2023/047-02

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT PRIMARY HEALTH CENTRE , KOLATHUR , METTUR (T.K.) 636 303	SRF No:	VCPL/2023/047
	Date of Calibration:	05-10-2023
	Date of issue:	06-10-2023
	Date of Receipt:	05-10-2023
	Next cal Date:	04-10-2024

### Details of Device Under Calibration

Description:	MICROPIPETTE	Make:	MICROLUX
Range:	( 10 to 50 ) $\mu$ l	Model/Type:	-
Resolution:	1 $\mu$ l	Condition of DUC	GOOD
Serial Number	-	Location:	PHC -LAB
ID number	KGPHC/MP/01	Cal. At:	LAB

### Environmental Conditions

Temperature	23.6 °C	<b>Calibration Method Used</b>	ISO 8655:2002 PART-6
Humidity	51 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

### RESULT OF CALIBRATION

S.NO	Duc Readings ( $\mu$ l )	Standard Reading Mean Value ( mg )	ERROR (mg)
1	10	9.93	0.07
2	25	24.89	0.11
3	50	49.84	0.16


### Remarks:

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

Calibrated By:

  
C.AKASH

(CALIBRATION ENGINEER)



Authorised by:

  
T. SUDHAHAR,  
Technical Manager

Vanavil Calibrations Pvt Ltd.

## Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com



## CALIBRATION CERTIFICATE

Certificate No: VCPL/CC/2023/047-03

Page no: 1 of 1

<b>Customer Name &amp; Address:</b> M/S GOVERNMENT PRIMARY HEALTH CENTRE , SATHYANAGAR , METTUR (T.K.) 636 303	SRF No:	VCPL/2023/047
	Date of Calibration:	05-10-2023
	Date of issue:	06-10-2023
	Date of Receipt:	05-10-2023
	Next cal Date:	04-10-2024

### Details of Device Under Calibration

Description:	MICROPIPETTE	Make:	MICROLUX
Range:	( 100 to 1000 ) µl	Model/Type:	-
Resolution:	1 µl	Condition of DUC	GOOD
Serial Number	-	Location:	PHC - LAB
ID number	KGPHC/MP/02	Cal. At:	LAB
<b>Environmental Conditions</b>		<b>Calibration Method Used</b>	
Temperature	23.4 °C	Calibration Method Used	ISO 8655:2002 PART-6
Humidity	50 % Rh	Cal procedure no:	-

### Master Details

SL no:	Description	ID No / SL No	Certificate No	Make/Model	Valid Till
1	ELECTRONIC BALANCE	16031172	E-150323-1	WINTECH	15-03-2024

### RESULT OF CALIBRATION

S.NO	Duc Readings ( µl )	Standard Reading Mean Value ( mg )	ERROR (mg)
1	100	99.91	0.09
2	500	499.87	0.13
3	1000	999.81	0.19


**Remarks:**

- 1.DUC is Defined as Device Under Calibration.
- 2.Calibration Frequency decided by the customer.
- 3.The above instrument is working in good condition.

**Calibrated By:**

  
**C.AKASH**

**(CALIBRATION ENGINEER)**



\*\*\* End of the Report \*\*\*

**Authorised by:**

  
**T. SUDHAKAR,**  
Technical Manager

**Vanavil Calibrations Pvt Ltd.**

## Vanavil Calibrations Pvt. Ltd

(A Unit of Rainbow Bio Medical Engineers)

S2, Rishi Maharaj Apartment, Indra Nagar, 250 Reddiyur Main Road, Narasothipatty, Salem - 636 004.

No. 96,97,98, Anadhamaya Apartments, 2<sup>nd</sup> Floor, KVR Flats, 2<sup>nd</sup> Cross Street, Thangam Nagar, Pallikaranai, Chennai - 600 100.

Cell: 99949 13872 / 99949 23872 Ph: 0427-7965376

✉ vanavilcalibrations@gmail.com 🌐 www.rainbowbiomedicalengineers.com | www.vanavilcalibrations.com