

## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13392	MEDICAL DEVICES	Page No:1 of 1
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00738-0001 & 07-10-2023	
Receipt Date	07-10-2023	
Calibration Date	07-10-2023	
Calibration Due	06-10-2024	

**Customer Name & Address**  
 GOVERNMENT PRIMARY HEALTH CENTRE,  
 THANTHONI-639007,KARUR DISTRICT.

Details of Device Under Calibration (DUC)			
Description	SEMI AUTO ANALYZER	Make & Model	CLINISYS & NA
Range	MULTI	Sr. No	3002810150873
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature:25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY

RESULTS			
S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ	Uncertainty in % ( ± )
		91	13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA	Uncertainty in % ( ± )
		165	7.6
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA	Uncertainty in % ( ± )
		225	7.2

**REMARKS**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By Authorised Signatory

  
 (Calibration Engineer)  
 K.SATHYAMOORTHY
 

 Technical Manager  
 C.SHANMUGARAJ
 

 Chief Executive

