



**CALIBRATION REPORT**

Report No : HBMS/CAL/23-24		Calibration Date : 28.09.2023
		Calibration Valid: 27.03.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		<b>M/s, UPGRADED PRIMARY HEALTH CENTRE, MAGUDANCHAVADI, SANKIRI, SALEM.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	RAJAS ENTERPRISES (INDIA)
ii	Model	----
iii	Serial No	5CP10AC
iv	Location	LABORATORY

**CALIBRATION WORKSHEET**

**Tabulation For – Centrifuge**

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured $\pm$ RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1502	2
4	2000	1999	-1

Acceptable deviation is  $\pm$  5 RPM it depends on Max Speed Range

