

**CALIBRATION CERTIFICATE**

Certificate No: SBS/CL/23/13396 Page No: 1 of 1

Customer Name & Address		SRF No.	SRF/23/00738-0005
GOVERNMENT PRIMARY HEALTH CENTRE, THANTHONI-639007,KARUR DISTRICT.		SRF Date	07-10-2023
		Date of Receipt	06-10-2023
		Date of Calibration	07-10-2023
		Due Date for Calibration	06-10-2024
		Issue Date	07-10-2023

Details of Unit Under Calibration			
Description	MICRO PIPETTE	Make	THERMO SCIENTIFIC
Range	100-1000µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition				Calibration Method Used	
Temperature (°C)	23.9	Humidity (%RH)	55	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08


Standard Used						
Sl. No.	Description	ID.No. / SI. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP122/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Z Factor: 1.00319


Result of Calibration in µl										
Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
1	100	99.95	99.96	99.94	99.98	99.96	99.96	-0.04	0.01	0.47
		99.96	99.97	99.98	99.97	99.95				
2	500	499.95	499.96	499.95	499.96	499.97	499.96	-0.04	0.01	0.47
		499.96	499.97	499.98	499.98	499.96				
3	1000	999.98	999.97	999.95	999.96	999.97	999.97	-0.03	0.01	0.47
		999.96	999.97	999.98	999.96	999.97				

**Remarks**

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- Equipment used for Calibration were calibrated & traceable to National & International Standards
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Calibration Liquid Used: Distilled or Deionized water conforming Grade3 as specified in ISO 3696.

Calibrated By,  
  
 (Calibration Engineer)  
 M.RAGUL



Authorised by:  
  
 Quality Manager/Chief Executive  
 G.SIVABALAN

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
 E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.