



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - JULY 2023



PC-1024

Lab Name Govt. PHC, Pannakkamudlu Lab No 12319
Constituent Group Chemistry III Date of Result Entered : 20/07/2023
PT item Lyophilized human serum based Date of Report Published : 07/08/2023

S.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1876	109.35	5.55	6.07	117 mg/dL	1.26	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1850	5.83	6.00	0.35	5.9 mg/dL	0.20	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1776	113.00	4.12	4.66	111 mg/dL	-0.43	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Pannakkamudlu
THE MEDICAL OFFICER, Kaveripattinam, Krishnagiri
Pannakkamudlu
TAMILNADU635106

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****