



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date: 28.09.2023
	Calibration valid: 27.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, UPGRADED PRIMARY HEALTH CENTRE, MAGUDANCHAVADI, SANKIRI, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	MICROLUX
ii Model	-----
iii Value	100µl-1000µl
iv Serial No	PHC/MGI/PIP/004

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	05-OCT-2023	04-OCT-2024

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.06	500.07	1000.17
2	100.09	500.05	1000.11
3	100.04	500.03	1000.09
4	100.08	500.06	1000.04
5	100.01	499.93	1000.03
6	100.00	500.09	1000.11

For Heart Beat Medical Systems,



 (Authorized signatory)