



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 15.09.2023
	Calibration valid : 14.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVERNMENT PRIMARY HEALTH CENTRE, KOOTTATHUPATTI , SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	P FACT
ii Model	-
iii Value	100µl-1000µl
iv Serial No	238649

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.02	500.12	1000.11
2	100.04	500.06	1000.04
3	100.05	500.03	1000.07
4	100.06	500.05	1000.14
5	100.02	500.03	1000.11
6	100.05	500.07	1000.04

For Heart Beat Medical Systems,


(Authorized signatory)





CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 15.09.2023
		Calibration valid : 14.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, GOVERNMENT PRIMARY HEALTH CENTRE, KOOTTATHUPATTI, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	P FACT
ii	Model	-
iii	Value	20 µl -100 µl
iv	Serial No	220154

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	20µl	50µl	100µl
1	20.03	50.03	100.05
2	20.04	50.05	100.05
3	20.02	50.02	100.04
4	20.02	50.04	100.03
5	20.00	50.02	100.04
6	20.02	50.07	100.02

For Heart Beat Medical Systems,



(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 15.09.2023 Calibration valid : 14.03.2024
1.1.CUSTOMER DETAILS:	
Name and Address of the organization	M/s, GOVERNMENT PRIMARY HEALTH CENTRE, KOOTTATHUPATTI , SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMO SCIENTIFIC
ii Model	FINEPITTE F3
iii Value	100µl-1000µl
iv Serial No	RW03149

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.01	500.09	1000.10
2	100.01	500.03	1000.04
3	100.05	500.05	1000.07
4	100.06	500.10	1000.15
5	100.02	500.05	1000.11
6	100.05	500.07	1000.04

For Heart Beat Medical Systems,

J. D. f
(Authorized signatory)





CALIBRATION REPORT

Report No : HBMS/CAL/23-24		Calibration Date : 15.09.2023
		Calibration valid : 14.03.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, GOVERNMENT PRIMARY HEALTH CENTRE, KOOTTATHUPATTI , SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		MICRO PIPETTE
i	Manufactured by	THERMO SCIENTIFIC
ii	Model	FINEPIPETTE F3
iii	Value	10 µl -100 µl
iv	Serial No	RW05774

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.04	50.02	100.05
2	10.04	50.05	100.05
3	10.03	50.02	100.04
4	10.02	50.04	100.03
5	10.00	50.02	100.04
6	10.02	50.07	100.02

For Heart Beat Medical Systems,

(Handwritten signature)

(Authorized signatory)



MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com