



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - MAY 2023



PC-1024

| | | | |
|-------------------|-------------------------------|----------------------------|------------|
| Lab Name | Govt. UPHC, Ansari Nagar | Lab No | 10778 |
| Constituent Group | Chemistry III | Date of Result Entered : | 18/05/2023 |
| PT item | Lyophilized human serum based | Date of Report Published : | 05/06/2023 |

| Sl.No | Analyte | Method / Principle Name | Analyzer Name | No. of Participants | CV | Participants CV | SD | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|---------------------|--------|-----------------|------|------------|-------|------|
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1834 | 180.88 | 5.19 | 9.39 | 203 mg/dL | 2.35 | 0.44 |
| 2 | CREATININE III | Jaffes Kinetic - Alkaline Picrate III | Robonik (Semi automatic / Fully Automatic) | 1854 | 1.76 | 7.48 | 0.13 | 1.6 mg/dL | -1.21 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1798 | 114.60 | 5.08 | 5.83 | 103 mg/dL | -1.99 | 0.27 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within ±1.01 to ±2.00 | Good. |
| Within ±2.01 to ±2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

LAB ADDRESS :

Govt. UPHC, Ansari Nagar
 MEDICAL OFFICER, ANSARI NAGAR
 Madurai-Corp
 TAMILNADU625010

Coordinator Contact Details:
 Email:clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

B. Shree
 MEDICAL OFFICER
 URBAN PRIMARY CENTRE
 ANSARI NAGAR,
 MADURAI CORPORATION