



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2023



10636

Name Govt. PHC, Pasuvanathanai

Lab No

19/06/2023

Instituent Group Chemistry III

Date of Result Entered :

04/07/2023

Item Lyophilized human serum based

Date of Report Published :

Sl.No	Analyte	Method	Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	
							CV	SD			
1	GLUCOSE III	GOD-POD III		Robonik (Semi automatic / Fully Automatic)	1823	286.63	5.17	14.82	296 mg/dL	0.63	0
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III		Robonik (Semi automatic / Fully Automatic)	1848	2.57	6.64	0.17	2.4 mg/dL	-0.99	0
3	CHOLESTROL III	CHOD-PAP III		Robonik (Semi automatic / Fully Automatic)	1804	113.73	4.46	5.07	116 mg/dL	0.45	0

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
 Govt. PHC, Pasuvanathanai
 CLINICAL LABORATORY , TUTICORIN District ..
 Pasuvanathanai
 TAMILNADU628718

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

[Signature]
 MEDICAL OFFICER
 GOVT. PRIMARY HEALTH CENTRE,
 PASUVANTHANAI
 KOVILPATTI - HUD