

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13586	MEDICAL DEVICES	Page No:1 of 1		
Issue Date	07-10-2023			
SRF No & Date	SRF/23/00773-0001 & 06-10-2023			
Receipt Date	06-10-203			
Calibration Date	06-10-2023			
Calibration Due	05-10-2024			
Customer Name & Address				
GOVERNMENT PRIMARY HEALTH CENTRE KARIYAPATTINAM-614806,NAGAPATTINAM DISTRICT.				
Details of Device Under Calibration (DUC)				
Description	SEMI AUTO ANALYZER	Make & Model : ROBONIK & PRIETEST TOUCH		
Range	MULTI	Sr. No : ATCD4371220RBK		
Resolution	MULTI	Identification No : NA		
DUC Condition	SATISFACTORY	Location : LABORATORY		
Environmental Conditions & Calibration Procedure Details				
Environmental Details	Temperature:25.6°C	Relative Humidity : 52% RH		
Calibration Procedure No	SBS/CP/MD/20	Calibration done at : ONSITE		
Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS


S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resitance >20MΩ	Measured values in MΩ 93	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 177	Uncertainty in % (±) 9.8
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 216	Uncertainty in % (±) 8.8

REMARKS


- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory


(Calibration Engineer)
K,SATHYAMOORTHY




Technical Manager
(C.SHANMUGARAJ)

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.