

CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/13453	MEDICAL DEVICES	Page No:1 of 1
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00752-0001 & 07-10-2023	
Receipt Date	07-10-2023	
Calibration Date	07-10-2023	
Calibration Due	06-10-2024	

Customer Name & Address
 GOVERNMENT PRIMARY HEALTH CENTRE,
 VISWANATHAPURI-639002,KARUR DISTRICT.

Details of Device Under Calibration (DUC)

Description : SEMI AUTO ANALYZER	Make & Model : ROBONIK & PRIETEST TOUCH
Range : MULTI	Sr. No : AT2170211RBK
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

Environmental Conditions & Calibration Procedure Details

Environmental Details	Temperature:25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

ELECTRICAL SAFETY

RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	Insulation Resisitance >20MΩ	Measured values in MΩ 87	Uncertainty in % (±) 13.92
2	Earth Leakage <5000µAfor B,BF,CF	Measured values in µA 234	Uncertainty in % (±) 8.5
3	Enclosure Leakage <500µAfor B,BF,CF	Measured values in µA 214	Uncertainty in % (±) 8.9


REMARKS

- This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- The results stated in this certificate relate only to the item calibrated.
- The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
- Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

K.S.H.
 (Calibration Engineer)
 K.SATHYAMOORTHY



CSL
 Technical Manager
 C.SHANMUGARAJ

Chief Executive