

		04110	RATION	CERTI	FICATE			
	CALIBRA			MEC	HANICAL	Page No:1 of 1		
CERTIFICATE NO: SBS/CL/23/13454  Issue Date SRF No & Date Receipt Date Calibration Date Calibration Due Customer Name & Address			SRF/23/0 07-10-20 07-10-20	07-10-2023 SRF/23/00752-0002 & 07-10-2023 07-10-2023 07-10-2023 06-10-2024				
COVER	RNMENT PRIMARY HEALTH	CENTRE,						
VISWA	NATHAPURI-639002, KARUR	DISTRICT	ils of Device Un	to - Callbrook	on (DUC)			
Environmental Details Temperature:			Sr. No Identification Location Sonditions & Star	Identification No Location  Identification No Relative Humidity  Outline to the state of the sta		AL MICRO & NA  NA  NA  LABORATORY  dure Details  54 % Rh  ONSITE		
Calibra	tion Procedure No	OBO/O/ /III	Reference Sta	andards Det	ails		11.17.476	
1	No Description M Digital Tachometer Lu		Make/ SI No: Lutron & Q6383	e/ SI No:		No Validity TR-A&S-2022-0036 03-10-2023		
S.No	DEVICE UNDER CALIBRATION READINGS	STANDARI	) INSTRUMENTS	NTS DEVIATION RPM		EXPANDED UNCERTAINTY (±		
			RPM			%		
	RPM 1000		999.7	0.3			4.2	
2	2000		1998	2			4.2	
3	3000		2998	2		4.2		

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By KHY Chennai (Calibration Engineer) 600 032 K.SATHYAMOORTHY

Chief Executive

**Authorised Signatory** 

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,