



**CHRISTIAN MEDICAL
COLLEGE**
DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE
SCHEME
EQAS YEARLY SUMMARY - 2022



PC-1024

Lab Name Govt. PHC, Masinaickenpatti Lab No 12577
Constituent Group Chemistry III
PT item Lyophilized human serum based Date of Report Published 20-02-2023

Group I : Laboratories who reported 10 to 12 months during 2022

Group II : Laboratories who reported 6 to 9 months during 2022

Group III : Laboratories who reported less than 6 months during 2022

Sl.No	Parameter	Rank Group	No of Labs	OMSDI	Grade
1	GLUCOSE III	Group I	1941	0.97	A
2	CREATININE III	Group I	1927	1.24	B
3	CHOLESTROL III	Group I	1923	1.05	B

Only first 2 Groups are Graded

OMSDI Range	Grade	Remarks
0.00 - 1.00	A	Excellent
1.01 - 2.00	B	Good
2.01 - 2.99	C	Satisfactory
Beyond ± 3.00	D	Unacceptable

LAB ADDRESS

Govt. PHC, Masinaickenpatti
masinaickenpatti phc
Masinaickenpatti
TAMILNADU636103

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College,
Vellore

Data in CMC EQAS reports is confidential
CMC EQAS does not subcontract any components
***** End of Report *****

R. Divya
மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்
மாசிநாயக்கன்பட்டி, சேலம்-636103



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2023



PC-1024

Lab Name Govt. PHC, Masinaickenpatti Lab No 12577
 Constituent Group Chemistry III Date of Result Entered : 20/09/2023
 PT item Lyophilized human serum based Date of Report Published : 05/10/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1870	310.86	4.82	14.98	310.9 mg/dL	0.00	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1852	1.29	8.14	0.10	1.32 mg/dL	0.29	0.00
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	113.62	3.80	4.32	116 mg/dL	0.55	0.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. PHC, Masinaickenpatti
 masinaickenpatti phc
 Masinaickenpatti
 TAMILNADU 636103

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

R. Ravi
 மருத்துவ அலுவலர்
 அரசு ஆரம்ப சுகாதார நிலையம்
 மாசிநாயக்கன்பட்டி, சேலம்-636103