



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - APRIL 2023



Lab Name: Govt. Block PHC, Kariapettinam
Constituent Group: Chemistry III
PT Item: Lympholized human serum based

Lab No: F112
Date of Result Entered: 18/04/2023
Date of Report Published: 06/05/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	CV	Participants CV	SD	Mean Value	SD	SI
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1832	118.45	8.04	7.21	119.8 mg/dL	8.05	0.34
2	CREATININE III	Jaffes Kojelic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1820	2.16	8.09	0.18	2.34 mg/dL	1.14	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1750	111.61	0.19	5.79	100.7 mg/dL	-1.88	0.28

SDI Range	Interpretation
Within ± 0.00 to ± 1.00	Excellent
Within ± 1.01 to ± 2.00	Good
Within ± 2.01 to ± 3.99	Accept with caution. Warning Signal
Beyond ± 4.0	Unacceptable performance. Action Signal

LAB ADDRESS :
Govt. Block PHC, Kariapettinam
The Block Medical officer, Kariapettinam, Nagapettinam District
Kariapettinam
TAMILNADU 614806

Coordinator Contact Details:
Email: clinqc@cmcvallore.ac.in
Contact Number: 0415-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
• Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****