



CALIBRATION REPORT	
Report No : HBMS/CAL/23-24	Calibration Date : 15.09.2023
	Calibration valid : 14.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	<b>M/s, GOVERNMENT PRIMARY HEALTH CENTRE, MASINAICKANPATTI , SALEM.</b>
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	CENTRIFUGE
i Manufactured by	REMI
ii Model	-
iii Serial No	AFLC-307

**CALIBRATION WORKSHEET**

LOCATION: LABORATORY

**Tabulation For – Centrifuge**

S.no	Unit under test IN Minutes	Observed speed tachometer in RPM	Deviation measured ± RPM
1	1000	1000	0
2	1200	1201	1
3	1500	1501	1
4	2000	2002	2

Acceptable deviation is ± 5 RPM it depends on Max Speed Range



MD Reg No: TN/CBE/MD42/00016



GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662



No.42/1, Navaindia Road, Peelamedu,

CBE – 04, hbmedicalsystem@gmail.com



S.no	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation: ± 0.5 seconds

### Testy equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer ( non – contact ) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

### Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by: 

