

CALIBRATION CERTIFICATE

DATE: 04.10.2023

NAME OF CUSTOMER:

MEDICAL OFFICER
 PRIMARY HEALTH CENTER
 ATHANUR PHC
 NAMAKKAL -DT

DESCRIPTION OF THE EQUIPMENT: CENTRIFUGE

MODEL NO: BDI-152

SL.NO: 2011842

COMPANY: SOLOKRAFTS INDUSTRIES

INSTRUMENT USED FOR CALIBRATION: DIGITAL TACHOMETER

MODEL NO:KUSAM-MECO&KM-2234BL

SL.NO: S1049947

CERTIFIED BY: MSIR INDIA PVT.

REGULATOR STEEPS	INSTRUMENT SPEED	TACHOMETER READING	ALLOWABLE ERROR	ERROR OBSERVED ±	RESULT
1.	700	713	±50	13	OK
3.	2100	2080	±50	20	OK
5.	3500	3485	±50	15	OK

NEXT CALIBRATION DUE ON: 03.10.2024

TESTED BY: G DINESH

APPROVED BY: Q.A

NOTE: 1) Calibration Points & due date is given as per customer request 2) This certificate refer only to particular item submitted for certificate 3) The calibration results reported are valid at the time of and under the stated conditions of the measurements 4) Any error in the certificate should

