

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PNC, Kallagattinam
Lab No 1112
Month June
Year 2023
Constituent Group Chemistry III

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

Date of Result Entered : 26/06/2023

Date of Report Published : 04/07/2023

| Sl No | Analyte | Method / Principle Name | Analyser Name | No of Participants | Std | Participants | | Your Value | SDI | U |
|-------|----------------|---------------------------------------|--|--------------------|--------|--------------|-------|------------|-------|------|
| | | | | | | CV | SD | | | |
| 1 | GLUCOSE III | GOD-POD III | Robonik (Semi automatic / Fully Automatic) | 1873 | 286.63 | 5.17 | 14.82 | 276 mg/dL | -0.72 | 0.68 |
| 2 | CREATININE III | Jaffe's Kinetic - Albumin Picrate III | Robonik (Semi automatic / Fully Automatic) | 1848 | 2.57 | 0.64 | 0.17 | 2.7 mg/dL | 0.76 | 0.01 |
| 3 | CHOLESTROL III | CHOD-PAP III | Robonik (Semi automatic / Fully Automatic) | 1804 | 113.73 | 4.46 | 0.07 | 117 mg/dL | 0.64 | 0.24 |

| SDI Range | Interpretation |
|-----------------------|--|
| Within -1.00 to +1.00 | Excellent. |
| Within +1.01 to +2.00 | Good. |
| Within +2.01 to +2.99 | Accept with caution. Warning Signal. |
| Beyond ±3.0 | Unacceptable performance. Action Signal. |

Links

| | | | |
|--------------------------------|---------------------------------|------------------------------------|--|
| Modify Address | Modify Methods | Submit Lab Results | View Lab Monthly Summary |
| Yearly Summary | Change Password | Home | Sign-out |