

CALIBRATION REPORT

Report No. HBMS/CAL/23-24	Calibration Date 20/09/2023
	Calibration valid 19/03/2024
I.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s. GOVT. UPGRADED PRIMARY HEALTH CENTRE, MANICKAMPALAYAM, THIRUCHENGODE, NAMAKKAL.
I.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	MICRO PIPETTE
i Manufactured by	THERMOSCIENTIFIC
ii Model	FINNPIPETTE F3
iii Value	10 µl - 100 µl
iv Serial No	RW05727

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Calibration test Report

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12



For Heart Beat Medical Systems,

(Authorized signatory)

MD Reg No: TN/CBE/MD42/00016

GSTIN: 33AAJFH1833M1ZB

99400 28355, 9751206106

88833 15662

No.42/1, Navaindia Road, Peelamedu,

CBE - 04, hbmedicalsystem@gmail.com