

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. Block PHC, Karampattanam	Details About Mutual Analysis
Lab Code	7112	Detail About Monthly Summary
Month	July	Detail about SDI
Year	2023	
Consultant Group	Chemistry III	

Click on the analyte to view Graphical Data

Date of Result Entered : 20/07/2023

Date of Report Published : 07/08/2023

Sl No.	Analyte	Method / Principle Name	Analyser Name	No of Participants	Participants			Your Value	SDI	U
					CV	CV	SD			
1	GLUCOSE III	GOD-POD III	Roborisk (Semi automatic / Fully Automatic)	1876	108.35	5.55	0.07	112 mg/dL	0.44	0.28
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Roborisk (Semi automatic / Fully Automatic)	1880	5.83	6.09	0.35	5.9 mg/dL	0.20	0.02
3	CHOLESTROL III	CHOD-PAP III	Roborisk (Semi automatic / Fully Automatic)	1778	113.09	4.12	4.88	118 mg/dL	1.07	0.22

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within +2.01 to +2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Address	Modify Methods	Submit Lab Results	View Lab Monthly Summary
Yearly Summary	Change Password	Home	Sign-out