

			CALIE	BRATIO	N CER	TIFIC	ATE		
CERT	IFICATE NO:	SBS/CL/2	23/13587			MECHAN	CAL	Page No:1 of 1	
Issue Date			07-10-2023						
SRF No & Date			SRF						
Receipt Date			06-10-203						
Calibration Date			06-10-2023						
Calibration Due			05-10-2024						
Custo	mer Name & Ad	dress							
GOVE	RNMENT PRIN	MARY HEALTH	CENTRE,						
KARI	YAPATTINAM-6	14806,NAGAP	ATTINAM DIS	STRICT.					
			Deta	ils of Device	Under Calib	ration (DU	C)		
Descr	iption :	CENTRIF	UGE	Mak	e & Model	:	REMI & R-8C		
Range	: •		3500 RPM	Sr. 1	No	:	ZHBN-07175		
Resol	ution :		1 RPM	Iden	tification No	:	NA		
DUC (Condition :	Satisfacto	ory	Loca	ation	:	LABORATORY		
			ronmental Co	onditions & S	tandard Ope	erating Pro	cedure Details		
Environmental Details Temperatur			Temperature	e: 25.5°C Relative Hui		midity 52%Rh			
Calibration Procedure No		No	SBS/CP/ML/04		Calibration done at		ONSITE	ONSITE	
				Reference	Standards D	etails			
S.No	Description				Make/ SI No:		e No	Validity	
1 Digital Tachometer			LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013		09-06-202		
CALIE	BRATION RESU	ILTS	-	Aller					

S.No	DEVICE UNDER CALIBRATION	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY	
	RPM	RPM	RPM	%	
1	500	498.9	1.1	4.2	
2	1500	1499.5	0.5	4.2	
3	3000	2999.5	0.5	4.2	

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

(Calibration Engineer) K.SATHYAMOORTHY



Authorised Signatory

(C.SIVABALAN) (Chief Executive)

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.