



## CALIBRATION CERTIFICATE

CERTIFICATE NO: SBS/CL/23/12889	MEDICAL DEVICES	Page No:1 of 1
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00640-0001 & 06-10-2023	
Receipt Date	06-10-2023	
Calibration Date	06-10-2023	
Calibration Due	05-10-2024	

**Customer Name & Address**  
GOVERNMENT PRIMARY HEALTH CENTRE,  
KADAMALAIKUNDU-625579, THENI DISTRICT.

Details of Device Under Calibration (DUC)			
Description	SEMI AUTO ANALYZER	Make & Model	ROBONIK & PRIETEST TOUCH
Range	MULTI	Sr. No	ATCD3641220RBK
Resolution	MULTI	Identification No	NA
DUC Condition	SATISFACTORY	Location	LABORATORY

Environmental Conditions & Calibration Procedure Details			
Environmental Details	Temperature: 25.6 ° C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/20	Calibration done at	ONSITE

Reference Standards Details				
S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

### ELECTRICAL SAFETY

#### RESULTS

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	<b>Insulation Resitance</b> >20MΩ	<b>Measured values in MΩ</b> 97	<b>Uncertainty in % ( ± )</b> 13.92
2	<b>Earth Leakage</b> <5000µAfor B,BF,CF	<b>Measured values in µA</b> 125	<b>Uncertainty in % ( ± )</b> 8.0
3	<b>Enclosure Leakage</b> <500µAfor B,BF,CF	<b>Measured values in µA</b> 235	<b>Uncertainty in % ( ± )</b> 7.2

#### REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
(Calibration Engineer)  
P.PRASANNA

Authorised Signatory

  
Technical Manager  
C.SHANMUGARAJ

Chief Executive



## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

CERTIFICATE NO:	SBS/CL/23/12890	MECHANICAL	Page No:1 of 1
Issue Date	07-10-2023		
SRF No & Date	SRF/23/00640-0002 & 06-10-2023		
Receipt Date	06-10-2023		
Calibration Date	06-10-2023		
Calibration Due	05-10-2024		

**Customer Name & Address**  
GOVERNMENT PRIMARY HEALTH CENTRE,  
KADAMALAIKUNDU-625579, THENI DISTRICT.

### Details of Device Under Calibration (DUC)

Description	: CENTRIFUGE	Make & Model	: M.C.DALAL&CO & NA
Range	: 3500 RPM	Sr. No	: MCD-419
Resolution	: 1 RPM	Identification No	: NA
DUC Condition	: Satisfactory	Location	: LABORATORY

### Environmental Conditions & Standard Operating Procedure Details

Environmental Details	Temperature: 24.3 °C	Relative Humidity	54 % Rh
Calibration Procedure No	SBS/CP/ML/04	Calibration done at	ONSITE

### Reference Standards Details

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Digital Tachometer	Lutron & DT-2238	JRPM-CCTR-A&S-2022-0036	03-10-2023

### CALIBRATION RESULTS

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY ( ± )
	RPM	RPM	RPM	%
1	1000	999.9	0.1	4.2
2	2000	1998	2	4.2
3	3000	2999	1	4.2

### REMARKS

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

  
(Calibration Engineer)  
P.PRASANNA



Authorised Signatory

  
 Quality Manager  
(C.SIVABALAN)

Chief Executive

## SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO: SBS/CL/23/12891</b>	<b>MEDICAL DEVICES</b>	<b>Page No:1 of 1</b>
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00640-0003 & 06-10-2023	
Receipt Date	06-10-2023	
Calibration Date	06-10-2023	
Calibration Due	05-10-2024	

**Customer Name & Address**GOVERNMENT PRIMARY HEALTH CENTRE,  
KADAMALAIKUNDU-625579, THENI DISTRICT.**Details of Device Under Calibration (DUC)**

Description	: ELECTRICAL SAFETY (MICROSCOPE)	Make & Model	: LABOMED & VISION 2000
Range	: MULTI	Sr. No	: CM/L943785
Resolution	: MULTI	Identification No	: NA
DUC Condition	: SATISFACTORY	Location	: LABORATORY

**Environmental Conditions & Calibration Procedure Details**

Environmental Details	Temperature: 25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/29	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

**ELECTRICAL SAFETY****RESULTS**

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	<b>Insulation Resitance</b>	<b>Measured values in MΩ</b>	<b>Uncertainty in % ( ± )</b>
	>20MΩ	97	13.92
2	<b>Earth Leakage</b>	<b>Measured values in μA</b>	<b>Uncertainty in % ( ± )</b>
	<5000μAfor B,BF,CF	125	12.3
3	<b>Enclosure Leakage</b>	<b>Measured values in μA</b>	<b>Uncertainty in % ( ± )</b>
	<500μAfor B,BF,CF	229	8.6

**REMARKS**

- 1.This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2.The user should determine the suitability of the instrument for its intended use.
- 3.The recalibration interval should be determined on the user requirement.
- 4.The results stated in this certificate relate only to the item calibrated.
- 5.The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00 .
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
(Calibration Engineer)  
P.PRASANNA

  
Technical Manager  
C.SHANMUGARAJ

Chief Executive

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.



## CALIBRATION CERTIFICATE

<b>CERTIFICATE NO: SBS/CL/23/12892</b>	<b>MEDICAL DEVICES</b>	<b>Page No:1 of 1</b>
Issue Date	07-10-2023	
SRF No & Date	SRF/23/00640-0004 & 06-10-2023	
Receipt Date	06-10-2023	
Calibration Date	06-10-2023	
Calibration Due	05-10-2024	

**Customer Name & Address**

GOVERNMENT PRIMARY HEALTH CENTRE,  
KADAMALAIKUNDU-625579, THENI DISTRICT.

**Details of Device Under Calibration (DUC)**

Description : HEAMOTOLOGY ANALYZER	Make & Model : SYSMEX & XP-100
Range : MULTI	Sr. No : B5903
Resolution : MULTI	Identification No : NA
DUC Condition : SATISFACTORY	Location : LABORATORY

**Environmental Conditions & Calibration Procedure Details**

Environmental Details	Temperature: 25.6°C	Relative Humidity	52% RH
Calibration Procedure No	SBS/CP/MD/24	Calibration done at	ONSITE

**Reference Standards Details**

S.No	Description	Make/ SI No:	Certificate No	Validity
1	Electrical Safety Analyser	Rigel Medical & 44L-1059	M-230809-16-4	10-08-2024

**ELECTRICAL SAFETY****RESULTS**

S.no	SPECIFICATION	MEASURED VALUES	EXPANDED UNCERTAINTY (±)
1	<b>Insulation Resistance</b> >20MΩ	<b>Measured values in MΩ</b> 88	<b>Uncertainty in % (±)</b> 13.92
2	<b>Earth Leakage</b> <5000μA for B,BF,CF	<b>Measured values in μA</b> 175	<b>Uncertainty in % (±)</b> 9.9
3	<b>Enclosure Leakage</b> <500μA for B,BF,CF	<b>Measured values in μA</b> 235	<b>Uncertainty in % (±)</b> 8.5

**REMARKS**

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

Authorised Signatory

  
(Calibration Engineer)  
P.PRASANNA



  
Technical Manager  
C.SHANMUGARAJ

Chief Executive

**SUNSHINE BIOMEDICAL SOLUTIONS**

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,  
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.