

2:00 PM

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. PHC, Palakurichi  
 Lab No 12386  
 Month March  
 Year 2023  
 Constituent Chemistry III  
 Group

Details About Robust Analysis  
 Detail About Monthly Summary  
 Detail about SDI

Click on the analyte to view Graphical Data

All Analyser Result

Print

Date of Result Entered : 20/03/2023

Date of Report Published : 04/04/2023

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	CV	Participants	Your Value	SDI	
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1878	91.48	7.46	6.82	92 mg/dL	0.08	0.31
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1851	1.33	9.08	0.12	1.35 mg/dL	0.17	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1761	110.93	5.23	5.80	114 mg/dL	0.53	0.28

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Access	Modify Methods	Submit Lab Results	View Lab Monthly Summary
View Summary	Usage Approved	Home	