



CALIBRATION REPORT		
Report No : HBMS/CAL/22-23		Calibration date : 22.09.2023
		Calibration valid: 21.03.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, PRIMARY HEALTH CARE CENTRE, KONERIPATTI ,SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		MICRO PIPETTE
I	Manufactured by	THERMOSCIENTIFIC
II	Model	FINNPIPETTE F3
III	Value	10-100 µl
IV	Serial No	QW10072
V	Location	LABORATORY

Details of the machine : machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	10µl	50µl	100µl
1	10.03	50.06	100.14
2	10.04	50.09	100.12
3	10.05	50.02	100.10
4	10.09	50.08	100.08
5	10.00	49.94	100.09
6	10.08	50.07	100.12

For Heart Beat Medical Systems,

(Authorized signatory)  
  





CALIBRATION REPORT		
Report No : HBMS/CAL/22-23		Calibration date : 22.09.2023
		Calibration valid: 21.03.2024
<b>1.1.CUSTOMER DETAILS:</b>		
Name and Address of the organization		M/s, PRIMARYHEALTH CARE CENTRE, KONERIPATTI, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>		
EQUIPMENT NAME		MICRO PIPETTE
I	Manufactured by	MICROLUX
II	Model	----
III	Value	100-1000 µl
IV	Serial No	PHC/KP/MP/002
V	Location	LABORATORY

Details of the machine : machine calibrated and checked output.

Number of measurement	Channel I (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	998.08
5	100.00	490.94	999.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

(Authorized signatory)



*[Handwritten Signature]*






CALIBRATION REPORT	
Report No : HBMS/CAL/22-23	Calibration date : 22.09.2023
	Calibration valid: 21.03.2024
<b>1.1.CUSTOMER DETAILS:</b>	
Name and Address of the organization	M/s, PRIMARYHEALTH CARE CENTRE, KONERIPATTI, SALEM.
<b>1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):</b>	
EQUIPMENT NAME	MICRO PIPETTE
I Manufactured by	THERMOSCIENTIFIC
II Model	FINNPIPETTE F3
III Value	100-1000 µl
IV Serial No	QW1481
V Location	LABORATORY

Details of the machine : machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	100µl	500µl	1000µl
1	100.03	500.06	1000.14
2	100.04	500.09	1000.12
3	100.05	500.02	1000.10
4	100.09	500.08	998.08
5	100.00	490.94	999.09
6	100.08	500.07	1000.12

For Heart Beat Medical Systems,

(Authorized signatory)



*R. [Signature]*

①

②

MD Reg No: TN/CBE/MD42/00016

99400 28355, 9751206106

No.42/1, Navaindia Road, Peelamedu,

GSTIN: 33AAJFH1833M1ZB

88833 15662

CBE – 04, hbmedicalsystem@gmail.com