

CALIBRATION REPORT	
Report No: HBMS/CAL/23-24	
Calibration Date: 23.09.2023	
Calibration Valid: 22.03.2024	
1.1. CUSTOMER DETAILS:	
Name and Address of the organization	M/s, UPGRADED PRIMARI HEALTH CENTRE, ARIYAPALAYAM, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):	
EQUIPMENT NAME	PIPETTE
I Manufactured by	THERMOSCIENTIFIC
II Model	FINNPIPETTE F3
III Value	5µl - 50µl
IV Serial No	UPHC/AM/PIP/004

Calibration standards used :

Master instrument	Model	Serial no	Date of calibration	Calibration due on
Electrical compact scale	SF-400C	SF400-2240	07-OCT-2022	06-OCT-2023

Details of the machine: machine calibrated and checked output.

Number of measurement	Channel 1 (Volume)		
	05µl	20µl	50µl
1	05.05	20.08	50.97
2	05.06	20.07	20.02
3	05.01	20.05	20.96
4	05.97	20.99	20.02
5	05.01	20.08	20.09
6	05.02	20.95	20.04

For Heart Beat Medical Systems,

(Authorized signatory)

