

CALIBRATION CERTIFICATE

Certificate No: SBS/CL/23/13324

Page No: 1 of 1

Customer Name & Address

GOVERNMENT PRIMARY HEALTH CENTRE,
NACHALUR-639110, KARUR DISTRICT.

SRF No.	SRF/23/00722-0004
SRF Date	07-10-2023
Date of Receipt	06-10-2023
Date of Calibration	07-10-2023
Due Date for Calibration	06-10-2024
Issue Date	07-10-2023

Details of Unit Under Calibration

Description	MICRO PIPETTE	Make	ERBA
Range	100-1000µl	Model	NA
Resolution	1µl	Material	PVC
Serial Number	NA	Operating Range	100-1000µl
ID Number	NA	Condition of UUC	Good
Cal. At	Mechanical Lab	Instrument Location	LABORATORY

Environmental Condition

Temperature (°C)	23.5	Humidity (%RH)	52	Calibration Method Used	National / International Standard	ISO 8655-6:2002
Atmospheric Pressure (mbar)	1006	Water Temperature (°C)	21.6	Cal Procedure No	SBS/CP/ML/08	

Standard Used

Sl. No.	Description	ID.No. / Sl. No.	Certificate No.	Make/Model	Traceability	Valid till
1	Electronic Weighing Balance	15112918	TVCSP22/12/2115-01	A&D & GH-252	National Standards	09-12-2023

Result of Calibration in µl

Z Factor: 1.00319


Sl. No.	Nominal Value	Observed Readings					Mean Value	Systematic Error	Random Error	Measurement Uncertainty (±)
		1	2	3	4	5				
1	100	99.94	99.96	99.94	99.92	99.93	99.95	-0.05	0.02	0.47
		99.94	99.95	99.96	99.97	99.98				
2	500	499.85	499.86	499.87	499.89	499.86	499.87	-0.13	0.02	0.47
		499.86	499.87	499.89	499.85	499.89				
3	1000	999.96	999.97	999.98	999.96	999.95	999.96	-0.04	0.01	0.47
		999.98	999.94	999.96	999.95	999.95				

Remarks

1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
2. The user should determine the suitability of the instrument for its intended use.
3. The recalibration interval should be determined on the user requirement.
4. The results stated in this certificate relate only to the item calibrated.
5. Equipment used for Calibration were calibrated & traceable to National & International Standards.
6. The indicated uncertainties are expanded uncertainty as defined for a confidence level of approximately 95% for a coverage factor k=2.00.
7. Calibration Liquid Used: Distilled or Deionized water conforming to Grade3 as specified in ISO 3696.

Calibrated By,

M. padmas
(Calibration Engineer)
MADHIBAN



Authorised by:

(Quality Manager/Chief Executive)
C.SIVABALAN

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087,
E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.