

		CAL	IBRA1	TION CE	RTIFIC	CAT	TE_			
CERTIFICATE N	o ses/o	L/23/13322			MECHA	IICAL	Р	age No:1 of 1		
Issue Date	, , , ,			07-10-2023						
SRF No & Date				SRF/23/00722-0002 & 07-10-2023						
Receipt Date				07-10-2023						
Calibration Date				07-10-2023						
Calibration Due				06-10-2024						
Customer Name	& Address									
GOVERNMENT		LTH CENTRE.								
NACHALUR-639	110 KARUR DIS	TRICT.								
		D	etails of De	vice Under Ca	libration (I	UC)				
Description	CENT	RIFUGE		Make & Model			B.D INSTRUMEN	ITATIONA & BDI-152		
Range	3	500 RPM		Sr. No			NA			
Resolution 1 RPM				Identification No			NA			
DUC Condition	Satisfa	ictory		Location			LABORATORY			
				s & Standard C	Operating F	roced	lure Details			
Environmental Details Temperature: 2							54 % Rh			
Calibration Procedure No SBS/CP/ML/			L/04	/04 Calibration		done at ONSITE				
			Refere	ence Standard	s Details					
S.No Description		Make/ S	Make/ SI No:		Certificate No		Validity			
1 Digital Tachometer		LINE SE	LINE SEIKI / 175-0034V		JRPM-CCTR-A&S-2023-0013		09-06-2024			

100	250	P1 1	101	4 1	450	UL	10

S.No	DEVICE UNDER CALIBRATION READINGS	STANDARD INSTRUMENTS	DEVIATION	EXPANDED UNCERTAINTY (±)
	RPM	RPM	RPM	%
1	1000	999.9	0.1	4.2
2	2000	1998	2	4.2
3	3000	2999	1	4.2

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- 2. The user should determine the suitability of the instrument for its intended use.
- The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor
- 6. Equipment used for Calibration were calibrated & traceable to National & International Standards.

Calibrated By

10 5/51 (Calibration Engineer) K.SATHYAMOORTHY



(C SIVABALAN)

Authorised Signatory

Chief Executive

SUNSHINE BIOMEDICAL SOLUTIONS

No:68, First Floor, Poomagal Main Road, Ekkattuthangal, Chennai - 600 032, Tel: 044 - 2225 2087, E-Mail: calibration@sunshinebiomedical.com, Website: www.sunshinebiomedical.com.