

				CAL	IBRA'	TIOI	CER	TIFICA	TE	
CERTIFICATE NO: SBS/CL/23/13619					MEDICAL DEVICES				Page No:1 of 1	
Issue Date SRF No & Date					07-10-2023 SRF/23/00782-0003 & 07-10-2023					
Receipt Date Calibration Date					07-10-2023 07-10-2023					
Calibration Due Customer Name & Address					06-10-2024					
				ITH CENTE ITHURALD	STRICT.	Device L	nder Calibr	ation (DUC)		
Description ELECTRICAL SAFETY (MICROSCOPE)				Is of Device Under Calibration Make & Model		ation (DUC)	LABOMED & VISI	ON 2000		
Range		1	MULTI			Sr. No		1	151285796	
Resolution		1	MULTI		Identification No			NA		
DUC Condition		1.6	SATISFACTORY			Location :		1	LABORATORY	
				Enviro	nmental Con-	ditions	& Calibratio	n Procedure	Details	
Environmental Details		tails	Temperature:25.6 * C			Relative Humidity		52% RH		
Calibr	ration Proce	dure No	SBS/CP/MD/29			Calibration done a		done at	t ONSITE	
					Refe	rence S	tandards De	etalls		
S,No	S.No Description			Make/ SI No:			Certificate No		Validity	
Electrical Safety Analyser			Rigel Medical & 44L-1059		-1050	M-230809-16-4		10-08-202		

RESULTS ELECTRICAL SAFETY								
1	Insulation Resisitance	Measured values in MΩ	Uncertainty in % (±)					
	>20MΩ	90	13.92					
2	Earth Leakage	Measured values in µA	Uncertainty in % (±)					
	<5000µAfor B,BF,CF	201	7.3					
3	Enclosure Leakage	Measured values in µA	Uncertainty in % (±)					
	<500µAfor B,BF,CF	215	7.3					

REMARKS

- 1. This Calibration certificate shall not be reproduced except in full, without written approval of the laboratory.
- The user should determine the suitability of the instrument for its intended use.
- 3. The recalibration interval should be determined on the user requirement.
- 4. The results stated in this certificate relate only to the item calibrated.
- 5. The indicated uncertainties are expanded uncertainty estimated for a confidence level of approximately 95% for a coverage factor k=2.00.
- Equipment used for Calibration were calibrated & traceable to National & International Standards.
 Calibrated By

Oundrained by

(Calibration Engineer) K.SATHYAMOORTHY Chennai Chenna

Technical Manager

Chief Executive

Authorised Signatory