

SSI : 2012/33/008/10489/LNSD
GSTIN : 33AGHPG4552G1ZO



NAVEEN TECHNOLOGIES

No.389, Dharsha Garden, Kandappa Colony, Annadanapatty, Salem - 2.
Cell : 94880 30384, 87787 93052 Email : mailnaveentech@gmail.com Web : www.naveentechnologies.in

CALIBRATION REPORT

Report No : NT/CAL/23-24/TDR/17		Calibration Date : 28.08.2023
		Calibration Valid: 27.02.2024
1.1.CUSTOMER DETAILS:		
Name and Address of the organization		M/s, PRIMARY HEALTH CENTRE, THEDAVUR, ATTUR, SALEM.
1.2. DESCRIPTION OF DEVICE UNDER TEST (DUT):		
EQUIPMENT NAME		CENTRIFUGE
i	Manufactured by	LABOTECH
ii	Model	-----
iii	Serial No	PHC/TDR/CEN/001

CALIBRATION WORKSHEET

LOCATION: LABORATORY

Tabulation For – Centrifuge

S.no	Unit under test in Minutes	Observed speed tachometer in RPM	Deviation measured \pm RPM	Combined uncertainty in RPM in % rdg
1	1000	1000	2	1.12
2	1200	1201	2	1.23
3	1500	1502	4	1.56
4	2000	1999	1	1.175

Acceptable deviation is \pm 5 RPM it depends on Max Speed Range



S.no	Unit under test in minutes	Standard reading Digital Timer in minutes	Deviation measured in minutes
1	05.0	05.1	0.1
2	10.0	10.2	0.2
3	15.0	15.1	0.1
4	20.0	20.2	0.2

Acceptance deviation: ± 0.5 seconds

Test equipment used

The calibration was performed by using Digital Multimeter & PT -100 sensor, Digital Timer (Stop Watch) and Digital Tachometer (non – contact) measured has been inspected in accordance with our inspection procedures to meets manufactures specification .

Remarks

This certificate may not be reproduced other than is full Except prior permission of laboratory.

Verified by : 



For Naveen Technologies,



(Authorised Signatory)